



Cellestis International Pty Ltd
1046A Dandenong Rd
Carnegie Victoria
Australia 3163
Ph: 61 3 9571 3500
Fax: 61 3 9571 3544
quantiferon@cellestis.com
www.cellestis.com
ACN 100 319 122

CERTIFICATE OF ANALYSIS

QuantiFERON® - TB Gold: High Altitude Mitogen tubes (Cat No.: T0593-0201)

Lot Number: 0593-60161

Date of Expiry: 2010-05

DESCRIPTION:	The Mitogen blood collection tubes are for use with the QuantiFERON®-TB Gold assay, which is a test of Cell Mediated Immunity for the detection of tuberculosis (TB) infection in humans. This simple and convenient blood test is designed to identify individuals who are infected with <i>Mycobacterium tuberculosis</i> complex (MTB) organisms. QuantiFERON®-TB Gold also provides an indication of the patient's general immune competence. Note: The blood collection tubes in this product are for use between altitudes of 1020m and 1875m only.
COMPOSITION:	The Mitogen tubes contain Phytohaemagglutinin-P (PHA-P) and heparin, which have been dried to the inner wall of the tubes.
PRESERVATIVE:	Not applicable
PRECAUTIONS:	Not applicable
STORAGE TEMPERATURE:	4°C to 25°C



Cellestis International Pty Ltd
1046A Dandenong Rd
Carnegie Victoria
Australia 3163
Ph: 61 3 9571 3500
Fax: 61 3 9571 3544
quantiferon@cellestis.com
www.cellestis.com
ACN 100 319 122

CERTIFICATE OF ANALYSIS

QuantiFERON® - TB Gold: High Altitude Mitogen tubes (Cat No.: T0593-0201)
Lot Number: 0593-60161
Date of Expiry: 2010-05

QUALITY CONTROL TESTING:

Mitogen tubes

Test	Criteria	Result
Visual examination	Clear coating visible on inside of tubes	Clear
Identity Test	Pass / Fail	Pass

NUMBER OF COMPONENTS: Components and Instruction leaflet

Component	Quantity per carton	Lot / Version Number	Expiry Date
Mitogen tubes	100	L020901	2010-05
Instruction Leaflet	1	L05900201K	Not applicable

This certifies that the results obtained for this batch of QuantiFERON®-TB Gold tubes conform to the quality specifications set by Cellestis Quality Control.

Dr. Stephen Jones
Chief Technical Officer, Cellestis Limited

Date: 16 June 2009